

# CERTIFICATE OF CALIBRATION

CERTIFICATE NO.: **SLM 28366**

**Equipment Description:** Sound Level Meter

**Manufacturer:** B & K

**Model No:** 2250      **Serial No:** 3007528

**Microphone Type:** 4952      **Serial No:** 3203504

**Preamplifier Type:** N/A      **Serial No:** N/A

**Filter Type:** 1/3 Octave      **Test No:** FILT 6129

**Comments:** All tests passed for class 1.  
(See over for details)

**Owner:** Envirosuite  
Level 3, 432 St Kilda Road  
Melbourne VIC 3004

**Ambient Pressure:** 997 hPa ±1.5 hPa

**Temperature:** 24 °C ±2° C      **Relative Humidity:** 55% ±5%

**Date of Calibration:** 24/11/2020      **Issue Date:** 24/11/2020

**Acu-Vib Test Procedure:** AVP10 (SLM) & AVP06 (Filters)

**CHECKED BY:** s47F      **AUTHORISED SIGNATURE:** s47F

Accredited for compliance with ISO/IEC 17025 - Calibration  
The results of the tests, calibration and/or measurements included in this document are traceable to Australian/national standards.



Accredited Lab. No. 9262  
Acoustic and Vibration  
Measurements



**HEAD OFFICE**  
Unit 14, 22 Hudson Ave. Castle Hill NSW 2154  
Tel: (02) 96808133 Fax: (02)96808233  
s47F  
web site: www.acu-vib.com.au